

# Optical Biometer AL-Scan US Edition




THE ART OF EYE CARE

## **Effortless Measurement of 6 Clinical Parameters in 10 Seconds**



NIDEK's solution is the state of the art optical biometer - the AL-Scan. In 10 seconds, six values for cataract surgery are measured:

- Axial length
- Corneal curvature radius
- Anterior chamber depth
- Central corneal thickness
- White-to-white distance
- Pupil size



#### **3-D Auto Tracking and Auto Shot**

The AL-Scan incorporates NIDEK's much acclaimed 3-D auto tracking and auto shot, enabling accurate measurement with ease and comfort. The 3-D auto tracking follows eye movements along the X-Y-Z directions to ensure accurate alignment of the eye. Once correct alignment is completed, the auto shot immediately captures the image and data.

# **Anterior Segment Observation with Scheimpflug Imaging** and Double Mire Ring Keratometry

The AL-Scan provides sectional lens image, pupil image, and reflected image of double mire rings projected onto the cornea.

The sectional lens image assists in the evaluation of the severity of the cataract. The pupil image assists in the assessment for multifocal IOL. The reflected image of mire rings assists in detecting an irregular corneal surface.



Sectional lens image (Scheimpflug image)



Pupil image




Reflected image of double mire rings

## **Optional Built-in Ultrasound Biometer**

In cases where the optical biometer cannot measure an eye with an extremely dense cataract, the AL-Scan provides an optional built-in ultrasound biometer, allowing measurement of virtually any cataractous eye with a combined model.

The AL-Scan requires no connection with an external ultrasound unit.



Ultrasound biometry



### **IOL Power Calculation and IOL Constants Optimization**

The IOL power is automatically calculated after measurement. Calculation of a personalized IOL constant improves postoperative accuracy.



#### **IOL** power calculation formula on AL-Scan

Nine IOL calculation formulas are incorporated in the AL-Scan. Once measurement is completed, the IOL power is automatically calculated using its own measured data.



Additional Barrett formulas available for the NAVIS-EX AL-Scan Viewer Barrett Universal II, Barrett True-K, Barrett Toric Calculator

# **AL-Scan Viewer** for NAVIS-EX

AL-Scan Viewer is software used for viewing and working with AL-Scan data via NAVIS-EX. This function enhances the capability of the AL-Scan with additional features and increases the efficiency of any clinic.



NAVIS-EX is an image filing software that enables data from the NIDEK diagnostic devices to be centralized in the NAVIS-EX database. It was initially developed for NIDEK's retinal products and has been expanded to network with the AL-Scan.

\* NAVIS-EX is optional software and is required for use of the AL-Scan Viewer.

#### **Data Management and IOL Power Calculations**


The large storage capacity of the NAVIS-EX database is available for review on the AL-Scan Viewer. The basic functions of the AL-Scan can also be performed with the AL-Scan Viewer including IOL power calculations and optimization of IOL constants.





#### Toric Lens Assist Function

Acquisition of multiple toric lens assist images allows selection of the optimal image for digitally marking the astigmatic axis. These images allow better surgical planning for accurate toric IOL alignment.





#### **Recalculation of Measured Values**

The AL-Scan Viewer allows recalculation of modified axial length, white-to-white, and pupil size data for accurate calculations.



#### **AL-Scan Specifications**

Optical measurement		
Axial length	Measurement range	14 to 40 mm
	Display increments	0.01 mm
	Measurement method	Low-coherence interferometry (LCI
Corneal curvature radius	Measurement range	5.00 to 13.00 mm
	Display increments	0.01 mm
Anterior chamber depth	Measurement range	1.5 to 6.5 mm
	Display increments	0.01 mm
Central corneal thickness	Measurement range	250 to 1,300 μm
	Display increments	1 μm
White-to-white distance	Measurement range	7 to 14 mm
	Display increments	0.1 mm
Pupil size	Measurement range	1 to 10 mm
	Display increments	0.1 mm
Ultrasonic measurement (optional	<u> </u>	
Axial length	Measurement range	12 to 40 mm
	Display increments	0.01 mm
Corneal thickness	Measurement range	200 to 1,300 μm
	Display increments	1 μm
IOL power calculation formula		·
Conventional Post-LASIK	Regression, Regression II, Formula/T Binkhorst, Hoffer Q,	
	Holladay 1, Haigis, Cam	
	Camellin-Calossi, Shammas PL	
Auto tracking	X-Y-Z directions	
Auto shot	Available	
Display	Tiltable 8.4-inch color LCD touch screen	
Printer	Thermal line printer with automatic paper cutter	
Interface	LAN, USB	
Power supply	AC 100 to 240 V	
	50/60 Hz	
Power consumption	100 VA	
Dimensions/Mass	283 (W) x 504 (D) x 457 (H) mm / 21 kg	
		11.1 (W) x 19.8 (D) x 18.0 (H)" / 46 lbs.
Al Con Violeor for NAVIS TV		
IOL calculation formula	Degracies Degra	UI Formula T Dinkhovet Us# O
		II, Formula/T, Binkhorst, Hoffer Q,
IOL calculation formula Conventional	Holladay 1, Haigis, Cam	nellin-Calossi, Barrett Universal II
IOL calculation formula Conventional Post-LASIK	Holladay 1, Haigis, Cam Camellin-Calossi, Sham	nellin-Calossi, Barrett Universal II mas-PL, Barrett True-K
IOL calculation formula Conventional Post-LASIK Toric calculator	Holladay 1, Haigis, Cam	nellin-Calossi, Barrett Universal II mas-PL, Barrett True-K
IOL calculation formula Conventional  Post-LASIK Toric calculator Additional features	Holladay 1, Haigis, Cam Camellin-Calossi, Sham Barrett Toric Calculator	nellin-Calossi, Barrett Universal II mas-PL, Barrett True-K
IOL calculation formula Conventional  Post-LASIK Toric calculator Additional features IOL registration	Holladay 1, Haigis, Cam Camellin-Calossi, Shami Barrett Toric Calculator Maximum data entry fo	nellin-Calossi, Barrett Universal II mas-PL, Barrett True-K or 100 IOLs
Conventional  Post-LASIK Toric calculator Additional features	Holladay 1, Haigis, Cam Camellin-Calossi, Sham Barrett Toric Calculator	nellin-Calossi, Barrett Universal II mas-PL, Barrett True-K or 100 IOLs



\* NAVIS-EX is optional software and is required for use of the AL-Scan Viewer.

Product/Model name: OPTICAL BIOMETER AL-Scan

Caution: U.S. Federal Law restricts this device to sale, distribution, and use by or on the order of a physician or other licensed eye care practitioner. Specifications may vary depending on circumstances in each country. Specifications and design are subject to change without notice.



NIDEK INC. 2040 Corporate Court San Jose, CA 95131 TEL: +1-800-223-9044 (US only) URL: http://usa.nidek.com

optimization

HEAD OFFICE (International Div.) 34-14 Maehama, Hiroishi-cho, Gamagori, Aichi 443-0038, JAPAN TEL: +81-533-67-8895 URL: www.nidek.com [Manufacturer]

TOKYO OFFICE (International Div.) 3F Sumitomo Fudosan Hongo Bldg, 3-22-5 Hongo, Bunkyo-ku, Tokyo 113-0033, JAPAN TEL:+813-5844-2641 URL: www.nidek.com

NIDEK S.A. Europarc, 13 rue Auguste Perret, 94042 Créteil, FRANCE TEL: +33-1-49 80 97 97 URL: www.nidek.fr NIDEK TECHNOLOGIES S.R.L. Via dell'Artigianato, 6/A, 35020 Albignasego (Padova), ITALY

TEL: +39 049 8629200/8626399 URL: www.nidektechnologies.it NIDEK (SHANGHAI) CO., LTD. Rm3205, Shanghai Multi Media Park, No.1027 Chang Ning Rd, Chang Ning District, Shanghai, CHINA 200050 TEL: +86 021-5212-7942 URL: www.nidek-china.cn NIDEK SINGAPORE PTE. LTD. 51 Changi Business Park Central 2, #06-14, The Signature 486066, SINGAPORE TEL: +65 6588 0389

URL: www.nidek.sg